



The advantages of spectral color characterization

Uwe Artmann

Image Engineering GmbH & Co KG




04.03.15

www.image-engineering.de



Color Image Workflow



The most simple form is a 3x3 matrix – the color correction matrix

$$\begin{bmatrix} R_{camera} \\ G_{camera} \\ B_{camera} \end{bmatrix} \begin{bmatrix} a_1 & a_2 & a_3 \\ b_1 & b_2 & b_3 \\ c_1 & c_2 & c_3 \end{bmatrix} = \begin{bmatrix} X \\ Y \\ Z \end{bmatrix}$$

Sensor-output

Color
correction
matrix

CIE XYZ


device
dependent

device
independent




More advanced system use Multidimensional Look-up-Tables (MLUT)

$$\begin{bmatrix} R_{camera} \\ G_{camera} \\ B_{camera} \end{bmatrix} MLUT = \begin{bmatrix} X \\ Y \\ Z \end{bmatrix}$$






Spectral response of II modules





Chromaticity ratio variation seen mid-way through production.



imageQuality basics

Color variation between individual and class calibration.



Individual Calibration



Average calibration (class calibration)





In general, there are two different methods to obtain the Camera Characterization matrix

- Chart-based method
- Spectra-based method



Chart-Based Method

RGB values created by the camera.



Camera
spectral sensitivity

unknown

measured





Chart-Based Method




X



X



= RGB




Lightsources
spectral distribution

Chart / Object
spectral reflection


Camera / human
spectral sensitivity



X



X



= XYZ



$$\begin{bmatrix} R_{camera} \\ G_{camera} \\ B_{camera} \end{bmatrix} \begin{bmatrix} a_1 & a_2 & a_3 \\ b_1 & b_2 & b_3 \\ c_1 & c_2 & c_3 \end{bmatrix} = \begin{bmatrix} X \\ Y \\ Z \end{bmatrix}$$

C

M


T

device
dependent

device
independent



Chart-Based Method



Idealized, assuming raw, linear, and noiseless where M is the estimator to be determined that transforms between camera responses C and tristimulus values T that minimizes colorimetric residual errors:

$$\| MC - T \|^2 \Rightarrow 0$$

First by linear estimation:

$$M = TC^t(CC^t)^{-1}$$

Generally followed by nonlinear optimization in perceptual coordinates (e.g. $L^*a^*b^*$, CIECAM02):

$$\begin{array}{c} f(T) \searrow \\ \text{Iterate } M: \| \Delta E \|^2 \Rightarrow 0 \\ f(MC) \nearrow \end{array}$$

Degrees of freedom can be reduced by constraining M to the scene adopted white point



Spectra-based method

$$R_{Camera} = k \cdot \int_{380nm}^{780nm} s(\lambda) \cdot r(\lambda) \cdot c_R(\lambda) \cdot d\lambda$$

$$G_{Camera} = k \cdot \int_{380nm}^{780nm} s(\lambda) \cdot r(\lambda) \cdot c_G(\lambda) \cdot d\lambda$$


$$B_{Camera} = k \cdot \int_{380nm}^{780nm} s(\lambda) \cdot r(\lambda) \cdot c_B(\lambda) \cdot d\lambda$$

with:

$s(\lambda)$ spectral distribution of the light source


$r(\lambda)$ spectral reflectance of the object

$c_x(\lambda)$ spectral sensitivity of the camera




Spectra-Based Method

RGB values created by the camera.



Spectra-Based Method



Lightsource
spectral distribution








Chart / Object
spectral reflection



Camera / human
spectral sensitivity

= RGB







= XYZ



Spectra-based method



← **$T = OIS$**
Training spectra colorimetric values



← **$C = RIS$**
Training spectra camera values

then: $T = MC$ as before
solve for M as illustrated previously




Chart:



- ✓ Simple, fast, and easy
- ✓ Minimal equipment required (target and colorimetric data)
- Valid transform domain is limited to the target gamut and capture conditions
- Multiple sets of target captures and pre-computed transforms required for each Original Scene adopted white point
- Chart colorants are not generally representative of likely Original Scene objects resulting in metameric errors

Limited to low-dimensionality transforms (matrix)

Spectra:




- ✓ Transform can be computed for any Original Scene adopted white point
- ✓ Training data is selected to be representative of Original Scene objects and radiation modes thereby minimizing metameric errors and optimizing for the likely use-cases
- ✓ Transforms are robust over a wide range of capture conditions and radiation modes
- ✓ Transforms are easily updated
- ✓ Suited to higher-dimensionality transforms (MLUT)
- Requires to measure camera spectral sensitivities





X-Rite ColorChecker SG



Color Chart
used in
Photography



04.03.15

www.image-engineering.de





Medical
Imaging has
different
challenges
for
ColorCharts






Small charts
are not
easily
produced



Chart:



Spectra:




Maybe not true for
medical imaging

- ✓ Simple, fast, and easy
- ✓ Minimal equipment required (target and colorimetric data)
- Valid transform domain is limited to the target gamut and capture conditions
- Multiple sets of target captures and pre-computed transforms required for each Original Scene adopted white point
- Chart colorants are not generally representative of likely Original Scene objects resulting in metameric errors

Limited to low-dimensionality transforms (matrix)

- ✓ Transform can be computed for any Original Scene adopted white point
- ✓ Training data is selected to be representative of Original Scene objects and radiation modes thereby minimizing metameric errors and optimizing for the likely use-cases
- ✓ Transforms are robust over a wide range of capture conditions and radiation modes
- ✓ Transforms are easily updated
- ✓ Suited to higher-dimensionality transforms (MLUT)
- Requires to measure camera spectral sensitivities






The maximum possible amount of patches is limited



Chart:



Spectra:



Strong limitations for medical imaging

- ✓ Simple, fast, and easy
- ✓ Minimal equipment required (target and colorimetric data)
- Valid transform domain is limited to the target gamut and capture conditions
- Multiple sets of target captures and pre-computed transforms required for each Original Scene adopted white point

Chart colorants are not generally representative of likely Original Scene objects resulting in metameric errors

Limited to low-dimensionality transforms (matrix)

- ✓ Transform can be computed for any Original Scene adopted white point
- ✓ Training data is selected to be representative of Original Scene objects and radiation modes thereby minimizing metameric errors and optimizing for the likely use-cases
- ✓ Transforms are robust over a wide range of capture conditions and radiation modes
- ✓ Transforms are easily updated
- ✓ Suited to higher-dimensionality transforms (MLUT)
- Requires to measure camera spectral sensitivities



Chart

vs. „real live“™






Figure 1: An example for a typical measurement setup to measure skin tones.



Figure 1: The image shows a typical measurement setup to measure skin tones.





The setup

element details

[add](#) [back](#)



| | | | |
|------------------|------|------------|-------|
| ID | 2440 | storage ID | ea146 |
| description | | category | |
| skin arm | | portrait | |
| illumination | | color | |
| daylight sunny | | skin asian | |
| brightness level | | | |
| highlight | | | |

original data:




white tile corrected:








Comparison of green tones





Principle component analysis of green colours



Upper row shows CCSG target patches, lower row in situ measured tones






Figure 7: The comparison of a variety of skin-tones as white tile corrected spectral radiances.



- Chart colorants are not generally representative of likely Original Scene
- This is especially true for medical applications as it is extremely hard to have e.g. “artificial skin”
- “in-situ” measured spectral radiances are a great help to optimize the color processing
- We are not aware that there is a database of “in-situ” measured spectral radiances of objects in the typical medical imaging use-cases
- Anyone interested in helping us to build one ?



Chart:




- ✓ Simple, fast, and easy
- ✓ Minimal equipment required (target and colorimetric data)
- Valid transform domain is limited to the target gamut and capture conditions
- Multiple sets of target captures and pre-computed transforms required for each Original Scene adopted white point

Chart colorants are not generally representative of likely Original Scene objects resulting in metameric errors

Limited to low-dimensionality transforms (matrix)

Spectra:



- ✓ Transform can be computed for any Original Scene adopted white point
- ✓ Training data is selected to be representative of Original Scene objects and radiation modes thereby minimizing metameric errors and optimizing for the likely use-cases
- ✓ Transforms are robust over a wide range of capture conditions and radiation modes
- ✓ Transforms are easily updated
- ✓ Suited to higher-dimensionality transforms (MLUT)
- Requires to measure camera spectral sensitivities



Spectral Sensitivity

Monochromator




Oriel MS257

Filter based



camspecs



camspecs
express

LED based



iQ-LED




| speed | cost | accuracy |
|----------------|--------|----------|
| slow | high | high |
| Medium to fast | medium | good |
| fast | medium | good |




Sample Hardware








Spectra of the 22 different LED Channels



- Measuring spectral sensitivities



- Individual Color-Calibration is important for best color reproduction
- The chart based calibration has various limitation, especially in the field of medical imaging
- The spectra based method is the better choice for use-case specific optimization
- Special “in-situ”-database for medical imaging would be very helpful





Image Engineering GmbH & Co. KG
Augustinusstraße 9d / 50226 Frechen / Germany
T +49 2234 995595 0 / F +49 2234 995595 10
www.image-engineering.de



04.03.15

www.image-engineering.de

